

EC Declaration of Conformity

<u>Name</u>	<u>Device Destination</u>	<u>Basic UDI-DI</u>
TABATABA® V2 (FF1006)	Expired CO detector for screening and prevention device in the battle against smoking addiction (for use in smoking cessation)	376025345FF1006T9

Medical devices conform to the following standards:

NF EN ISO 13485 :2016 : Medical Device – Quality Management System
 NF ISO 2859-1 :2000 : Sampling procedures for inspection by attributes — Part 1: Sampling schemes indexed by acceptance quality limit (AQL) for lot-by-lot inspection
 EN 60601-1:2006/A1:2013: Medical electrical equipment – Part 1: General requirements for basic safety
 EN 60601-1-2:2015: Medical electrical equipment - Part 1-2: General requirements for basic safety - Collateral standard: Electromagnetic compatibility - Requirements and tests
 IEC 60601-1-6:2007/AC: 2010: Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability
 EN 62366-1:2015: Medical devices - Application of usability engineering to medical devices
 EN ISO 10993-1:2009/AC: 2010: Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process.
 EN ISO 10993-5:2009: Biological evaluation of medical devices - Partie 5 : Tests for in vitro cytotoxicity
 EN ISO 10993-10:2010: Biological evaluation of medical devices - Partie 10 : Tests for irritation and skin sensitization
 NF EN ISO 14971:2019: Medical devices - Application of risk management to medical devices
 NF EN 62304/A1:2018 : Medical device software -- Software life cycle processes
 NF EN 1041 + A1: 2013: Information supplied by the manufacturer of medical devices
 EN ISO 15223-1:2016: Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General

I the undersigned, Dounia Benbachir, Quality Manager of the FIM MEDICAL company located at 51 rue Antoine Primat Villeurbanne - FRANCE assure and declare that the medical devices listed above belong to class Im (Rule 12) and satisfy the provisions of annex I (Essential Requirements), annex VI (Product Quality Assurance) and annex VII (Evaluation of Tabataba V2, FF1006DTR100 technical file) of the 93/42/EEC directive and its local adaptation (Book V of the public health code).

The devices described above are covered by the EC Certificate n° 27673 delivered by LNE/G-MED, 1 rue Gaston Boissier, 75724 Paris Cedex 15.

Villeurbanne, 12/07/2021,

D.BENBACHIR

Responsable Qualité



CE
0459