

## EC Declaration of Conformity

<u>Name</u>	<u>Matter</u>	<u>Device Description</u>
Nose-Clip (FF1031)	Polypropylene	Single-use nose-clip

Medical devices conform to the following standards:

EN ISO 10993-1:2009/AC: 2010: Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process.  
 EN ISO 10993-5:2009: Biological evaluation of medical devices - Partie 5 : Tests for in vitro cytotoxicity  
 EN ISO 10993-10:2010: Biological evaluation of medical devices - Partie 10 : Tests for irritation and skin sensitization  
 NF EN ISO 14971:2013: Medical devices - Application of risk management to medical devices  
 NF EN 1041 + A1: 2013: Information supplied by the manufacturer of medical devices  
 EN ISO 15223-1:2016: Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General

I the undersigned, Marie-Ange DEREI, President of the FIM MEDICAL company located at 51 rue Antoine Primat Villeurbanne - FRANCE assure and declare that the medical devices listed above belong to class I (Rule 12) and satisfy the provisions of annex VI (Product Quality Assurance) of the 93/42/EEC.

The devices described above are covered by the EC Certificate n° 27671 delivered by LNE/G-MED, 1 rue Gaston Boissier, 75724 Paris Cedex 15.

Villeurbanne, 8 August 2018,  
 Marie-Ange DEREI  
 Président


